

Aktive vibration isolation

We have two stages: R1 & clock laser are on one stage and R2 is on another stage.

- Manual: <hi #fff200>TODO</hi>
- three modi available: **active**, **transportable**, **off**
- Information on **active** modus:
- Information on **transportable** modus:
- Information on **off** modus:



Important in **active modus**: The “four bars” VS, HS, VA, HA should be in the upper third of the range for an adequate control range.



Movements in the room have influence of the vibration isolation!

From:

<https://iqwiki.iqo.uni-hannover.de/> - IQwiki

Permanent link:

https://iqwiki.iqo.uni-hannover.de/doku.php?id=groups:mg:aktiv_vibration_isolation&rev=1475518893

Last update: **2016/10/03 18:21**

