

Phase Noise Measurement

- We measured the phase noise with FSWP from "Rohde und Schwarz". Manual: fswp_dat-sw_en_3607-2090-22_v0400.pdf
- Information about the measurements is in the ResLab book on the pages 134 - 136 (date: 18.08.2016).
- measurement:
 - each device was measured directly behind the output, except it's mentioned
 - use SMA-cables instead of BNC-cables if possible, tighten the SMA-plug with an open-end wrench, use golden SMA-I-Connection (better quality)
 - don't wiggle the cable → results in noise in the range of 1Hz - 200Hz
 - let the cable swing off bevor starting the measurement
 - check before the measurement, if the dBm number isn't too high for the device
- analysis method:
 - Program "phasenoise_combined.nb" is located in the AFS
 - Copy the program into the folder with the data, which needs to be analyzed and start the program
- physical units:
 - dB: epresses the ratio of two values of a physical quantity
 - dBc (decibels relative to the carrier): power ratio of a signal to a carrier signal, expressed in decibels(dB)
 - dBm: abbreviation for the power ratio in decibels (dB) of the measured power referenced to one milliwatt

<hi #ff7f27> - Wie hatten wir gemessen? Was verwendet?

1. Klaus Zettel verwenden



2. Wie hatten wir die Datei gespeichert?

3. Was is dB, dBm, dBc und wie hängt W damit zusammen?


4. Frequenzen und dBm-Zahl zu jeder Messung angeben. Gibt es noch andere wichtige Daten in der Datei?</hi>



Beim nächsten mal Messen: klären wie gut die Geräte bei anderen Frequenzen funktionieren? Kabel überprüfen, Frequenzen hinter Originalkabel bis hin zum zum Beispiel AOM prüfen, Steckverbinder wie BNC-I überprüfen

AtomLab


atom_lab.pdf



- 10mhz atomslab.DAT: 9.999.999,686110 Hz [18.08.2016]
- aom rf atomlab nach abschwaecher.DAT: 87.169.997,264110 Hz [22.08.2016]
- aom rf atomlab nach kabelende.DAT: 87.169.997,256448 Hz [22.08.2016]
- aom rf atomlab nach verstaerker.DAT: 87.169.997,268579 Hz [22.08.2016]
- wutz.DAT: 104.673.996,567979 Hz [22.08.2016]

Resonator 1


r1.pdf



- clocksynth.DAT: 87.169.997,232243 Hz [18.08.2016]

Rb clock


rb_clock.pdf



- rb 10 mhz links.DAT: 9.999.999,682126 Hz [22.08.2016]
- rb 10 mhz rechts.DAT: 9.99.999,681785 Hz [22.08.2016]
- rb 5 mhz 2.DAT: 4.999.999,836086 Hz [22.08.2016]
- rb 5 mhz.DAT: 4.999.999,850747 Hz [22.08.2016]

fiber stabilization


fiber_stab.pdf



- 80 mhz.DAT: 79.110.997,503162 Hz [18.08.2016]
- synth. fiber stab. ext. mod.DAT: 79.110.937,461429 Hz [22.08.2016]
- synth.fiber stab. refext..DAT: 79.110.997,261913 Hz [22.08.2016]

Synthesizer von Q-Port

qport.pdf




- qport internal.DAT: 987.732.665,994588 Hz [22.08.2016]

- qport 10mhz ext. direkt stab. auf ocxo.DAT: 987.733.301,133802 Hz [22.08.2016]
- qport 10mhz ext.internal.DAT: 987.733.301,151171 Hz [22.08.2016]

Resonator 2


r2.pdf



- eom r2 sznth.DAT: 16.200.005,142169 Hz [24.08.2016]
- synth. r2 aom refext..DAT: 77.099.997,538021 Hz [22.08.2016]
- synth. r2 aom refext. richtig.DAT: 77.099.997,514594 Hz [22.08.2016]

Resonator 1 and 2


r1-r2.pdf



- synt. r1-r2 10mhz ext. direkt.DAT: 340.999.989,006932 Hz [22.08.2016]
- synz. r1-r2 10mhz ext. indirekt.DAT: 340.99.989,008668 Hz [22.08.2016]
- synt. r1-r2 internal.DAT: 341.000.016,201303 Hz [22.08.2016]

rep rate


rep_rate.pdf



- synth. rep rate 10mhz ext. direkt.DAT: 1.000.000.248,806976 Hz [22.08.2016]
- synth. rep rate 10mhz ext. direkt with 10mhz with real cable.DAT: 1.000.000.249,473547 Hz [22.08.2016]
- synth. rep rate 10mhz ext. indirekt.DAT: 1.000.000.248,770376 Hz [22.08.2016]
- synth. rep rate 10mhz ext. indirekt with real cable.DAT: 999.997.802,893996 Hz [22.08.2016]
- synth. rep rate 10mhz ext. indirekt with real cable with 10mhz.DAT: 1.000.000.249,567324 Hz [22.08.2016]
- synth. rep rate internal.DAT: 1.000.000.282,423948 Hz [22.08.2016]

PDH


pdh.pdf



- pdhlock2.DAT: 13.810.003,192828 Hz [23.08.2016]
- pdhlock.DAT: 13.817.786,393137 Hz [23.08.2016]
- pdhlock int.DAT: 13.810.004,243357 Hz [23.08.2016]
- pdhunlocked.DAT: 13.60.299,455637 Hz [23.08.2016]

10 MHz distribution


10_mhz_distribution.pdf



- 10mhz atomslab.DAT: 9.999.999,686110 Hz [18.08.2016]
- 10mhz k&k verteiler.DAT: 9.999.999,651353 Hz [18.08.2016]
- 10mhz ocxo stab indirect synth 2.DAT: 9.999.999,673212 Hz [22.08.2016]
- 10mhz ocxo stab synth.DAT: 9.999.999,673767 Hz [22.08.2016]
- ocxo 3.DAT: 4.999.999,834145 Hz [22.08.2016]

10 MHz synthesizer

10_mhz_synthesizer.pdf




- 10mhz internal synth 2.DAT: 10.000.000,225102 Hz [22.08.2016]
- 10mhz ocxo stab indirect synth 1.DAT: 9.999.999,677410 Hz [22.08.2016]
- 10mhz ocxo stab synth 2.DAT: 9.999.999,672932 Hz [22.08.2016]
- 10mhz ocxo stab synth 3.DAT: 9.999.999,672579 Hz [22.08.2016]

- 10mhz open synth 2.DAT: 9.999.474,907956 Hz [22.08.2016]
- 10mhz selbst.DAT: 9.99.999,650255 Hz [18.08.2016]
- 10mhz selbst no others.DAT: 9.999.999,640838 Hz [18.08.2016]

10MHz OCXO doubling

5MHz moving cable test


5mhz_moving_cable_test.pdf



- 5mhz moving cable test.DAT: 4.999.999,822782 Hz [18.08.2016]
- 5mhz without moving.DAT: 4.999.999,825716 Hz [18.08.2016]

5MHz to 10MHz OCXO


5mhz-ocxo_the_way_to_10mhz.pdf



- 5mhz + amp.DAT: 5.000.000,518799 Hz [18.08.2016]
- 5mhz.DAT: 4.999.999,825716 Hz [18.08.2016]
- 5mhz + doubler + amp.DAT: 9.999.999,658937 Hz [18.08.2016]
- 5mhz + doubler.DAT: 9.999.999,651068 Hz [18.08.2016]
- 5mhz + splitter.DAT: 4.999.999,820720 Hz [18.08.2016]

5MHz to 10MHz OCXO again


5mhz_to_10mhz_ocxo_again.pdf



- after ocxo and doubler.DAT: 10.000.000,353714 Hz [22.08.2016]
- after ocxo.DAT: 4.999.999,838434 Hz [22.08.2016]
- ocxo 3.DAT: 4.999.999,834145;Hz [22.08.2016]

5MHz OCXO with/without original power supply


5mhz_ocxo_with_without_original_power_supply.pdf



- 5mhzrg58smaocxo.DAT: 4.999.999,836230 Hz [18.08.2016]
- 5mhzrg58smaocxo w k+k supply ext supply.DAT: 4.999.999,830501 Hz [18.08.2016]
- 5mhzrg58smaocxo wo k+k supply ext supply.DAT: 4.999.999,824645 Hz [18.08.2016]

10MHz amplifier

10_mhz_amplifier.pdf



- 10mhz amp.DAT: 9.999.999,652707 Hz [18.08.2016]
- 10mhz amp mass_002.DAT: 9.999.999,686110 Hz [19.08.2016]
- 10mhz amp mass.DAT: 9.999.999,651541 Hz [18.08.2016]

Last
update:
2016/10/28 groups:mg:phase_noise_measurement https://iqwiki.iqo.uni-hannover.de/doku.php?id=groups:mg:phase_noise_measurement&rev=1477644750
08:52

From:
<https://iqwiki.iqo.uni-hannover.de/> - IQwiki

Permanent link:
https://iqwiki.iqo.uni-hannover.de/doku.php?id=groups:mg:phase_noise_measurement&rev=1477644750

Last update: **2016/10/28 08:52**

