

Spacer

Company: Menlo Systemes

- Material/Grade: 7972/Premium/Special
- Diameter: 92 mm (Measured: **91.75mm - 91.77mm**)
- Length: 485 mm

ULE Stabbearbeitung

Company: Siliciumbearbeitung Andrea Holm GmbH

- Rundschleifen
 - DM 91,7 (+/-0,1) mm
- Kernbohrung
 - DM: 13 (+/-0,2) mm
- Seitliche Bohrung
 - DM: 4,2 (+/-0,2) mm

Notes vom Rohblock

Item is oversized cores to yield the above sizes

- Finish: Waterjet or saw cut on all surfaces
- Tolerances: +12 mm/-0 on all dimensions
- Special CTE Selection: Crossover temperature $\geq 28^{\circ}\text{C}$ (in einer anderen mail, die ich fand stand: 28-35°C)

2D drawings

- Resonatorkörper
 - spacer_mikko
- Spacer mit Schutzkappen

From:

<https://iqwiki.iqo.uni-hannover.de/> - IQwiki

Permanent link:

https://iqwiki.iqo.uni-hannover.de/doku.php?id=groups:mg:project_ptb-cavity:spacer

Last update: **2017/08/08 08:17**



