

Substrates

Company: Coastline Optics, Inc.

Quotation:

q14-2520_university_of_hannover.pdf

25.4mm Ø, Plano/Plano Fused Silica Substrates

- Dimensions
 - Diameter: 25.4mm +0.0/-0.1 mm
 - Thickness: 6.35mm +/-0.2mm
- Configuration: Plano/Plano
- Side 1, central 5mm Ø area:
 - Surface Quality: < 10-5 - Scratch-Dig - per Mil- PRF-13830B.
 - Surface Roughness: < 1Angstrom. RMS asmeasured onour Zygo5500optical profiler.
 - Surface Figure: < 0.10wave P-V @ 633nm.
- Side 1, annular area between 18.5mm Ø and 25.4mm Ø:
 - Surface Quality: < 20-10 scratch dig, suitable for optical contact.
 - Surface Figure: < 0.25wave P-V @ 633nm
- Side 2, central 80% area:
 - Surface quality: 20-10scratch-dig
 - Surface figure: <0.10wave P-V @ 633nm
- Bevels: 0.020" - 0.040" x 45degree facewidthnominal
- Wedge: 10arc minutes
- Material: Corning 79800A Fused Silica
- We bought 4.

25.4mm Ø, 1´Meter Concave ROC/Plano Fused Silica Substrates

- Same as "25.4mm Ø, Plano/Plano Fused Silica Substrates"
- Configuration: Concave/Plano
- Side 1: -100cm+/-1cm concave ROC
- Material: Corning 79800A Fused Silica

- We bought 4.

IBS AR coating

- IBS AR coat Side 2 of the above Item1and2substrates for $R < 0.10\%$ @ 1550nm

From:
<https://iqwiki.iqo.uni-hannover.de/> - IQwiki

Permanent link:
https://iqwiki.iqo.uni-hannover.de/doku.php?id=groups:mg:project_ptb-cavity:substrates

Last update: **2017/05/11 09:24**

