



Draufsicht




Unb'sicht





120

120






Durchgangsloch
mit
Ø 14
Senkopt

Platte 3



10 / 20 125 /





Kühlkörper
 $1,3 \text{ K/W}$
 $100 \times 100 \times 40$

PT100 NB-PTCO-059
 Klasse AA
 $-30^\circ\text{C} - +200^\circ\text{C}$
 $2,0 \times 2,3 \text{ mm} \pm 0,2$
 $\times 1,1 \times 0,15$

NTC - kalibrier
 (KUK)
 $\varnothing 2 \text{ mm}$
 Höhe 8 mm

NTC - Vakuum
 (RS: 769-1937)
 $\varnothing 2,3 \pm 0,2 \text{ mm}$
 Höhe: $4,1 \pm 0,5 \text{ mm}$
 $\pm 1-3\%$ $\nearrow +3 \text{ mm}$