



Checklist Resonator Lab (D215)

1. a) Check **IGP Controller** for the resonators R1 + R2,
Current < 2 mA (both should show „HV ON“ (green LED on)
b) Check both **USV for IGP**



2. a) Check **Meerstetter TEC** for R1 + ECDL (green LED on)
b) Check both **USV for Meerstetter TEC**



3. a) Temperature 21°C?
b) See water?
c) Smell burning/see smoke?
d) Computer are running?

In emergency, please call:

Steffen Sauer:	0172 2068556
Klaus Zipfel:	0160 1602392
Nandan Jha:	0151 6357240
Dominika Fim:	0152 53424762